



GIMA

PROFESSIONAL MEDICAL PRODUCTS

GB - USER MANUAL FOR NON-SURGICAL INSTRUMENTS

These instruments are intended to be used by medical practitioners who are specially trained on how to use and care of them. The incorrect use, poor or inappropriate maintenance can rapidly lead to deterioration of the instruments. The first time, and after every use it is recommended to clean, dry the instruments. The instruments must always be cleaned prior to sterilisation. For automated cleaning use only washing equipment with approved and certified detergents. For manual cleaning use approved and certified detergents, brush and running water. Always follow instructions on how to use the detergent; clean the instruments both when open and closed; rinse for 3 minutes and check that water also enters and exits the blind holes several times. Use completely demineralised water in the final rinse phase. Instruments that are not dried could suffer damage by corrosion. Always dry the instruments. Also make sure that the product does not come into contact with acids or other aggressive disinfectants that could corrode it.

Instruments that show signs of corrosion must be discarded immediately. Always perform a visual inspection for damage or signs of wear: sharp edges must be free of dents and with continuous edges; there should be no distortion of instruments with long parts; the instruments that are part of a larger assembly, must be checked together with other assembly components; always check the rotating movement of hinges, which must not have excessive play; always check that the locking systems are working. Whilst the instruments are under warranty, repairs and/or replacement of all defective parts, for reasons that are accepted by the manufacturer, are free, with the exception of costs for labour, travel, transport, packing, etc. Damage caused by improper use of the product is not covered by the warranty.

GIMA WARRANTY TERMS









The Gima 12-month standard B2B warranty applies.

WARNINGS

All serious accidents concerning the medical device supplied by us must be reported to the manufacturer and competent authority of the member state where your registered office is located.

Simboli / Symbols / Symboles / Symbole / Simbolos / Simbolos / Σύμβολα / حرف

	<p>IT Fabbricante GB Manufacturer FR Fabricant ES Fabricante PT Fabricante DE Hersteller GR Παραγωγός</p> <p style="text-align: right;">SA</p>
	<p>IT Data di fabbricazione GB Date of manufacture FR Date de fabrication ES Fecha de fabricación PT Data de fabrico DE Herstellungsdatum GR Ημερομηνία παραγωγής</p> <p style="text-align: right;">SA</p>
	<p>IT Non sterile GB Non-sterile FR Pas stérile ES No estéril PT Não estéril DE Nicht steril GR Όχι αποστειρωμένο</p> <p style="text-align: right;">SA ليس معقم</p>

	<p>IT Attenzione: Leggere e seguire attentamente le istruzioni (avvertenze) per l'uso GB Caution: read instructions (warnings) carefully FR Attention: lisez attentivement les instructions (avertissements) ES Precaución: lea las instrucciones (advertencias) cuidadosamente PT Cuidado: leia as instruções (avisos) cuidadosamente DE Achtung: Anweisungen (Warnungen) sorgfältig lesen GR Προσοχή: διαβάστε προσεκτικά τις οδηγίες (ενστάσεις)</p> <p style="text-align: right;">SA الحذر: قراءة التعليمات (التحذيرات) بعناية</p>
	<p>IT Leggere le istruzioni per l'uso GB Consult instructions for use FR Consulter les instructions d'utilisation ES Consultar las instrucciones de uso PT Consulte as instruções de uso DE Gebrauchsanweisung beachten GR Διαβάστε προσεκτικά τις οδηγίες χρήσης</p> <p style="text-align: right;">SA</p>
	<p>IT Conservare al riparo dalla luce solare GB Keep away from sunlight FR À conserver à l'abri de la lumière du soleil ES Conservar al amparo de la luz solar PT Guardar ao abrigo da luz solar DE Vor Sonneneinstrahlung geschützt lagern GR Κρατήστε το μακριά από ηλιακή ακτινοβολία</p> <p style="text-align: right;">SA يحفظ بعيدا عن أشعة الشمس</p>
	<p>IT Conservare in luogo fresco ed asciutto GB Keep in a cool, dry place FR À conserver dans un endroit frais et sec ES Conservar en un lugar fresco y seco PT Armazenar em local fresco e seco DE An einem kühlen und trockenen Ort lagern GR Διατηρείται σε δροσερό και στεγνό περιβάλλον</p> <p style="text-align: right;">SA</p>
	<p>IT Codice prodotto GB Product code FR Code produit ES Código producto PT Código produto DE Erzeugniscode GR Κωδικός προϊόντος</p> <p style="text-align: right;">SA كود المنتج</p>
	<p>IT Numero di lotto GB Lot number FR Numéro de lot ES Número de lote PT Número de lote DE Chargennummer GR Αριθμός παρτίδας</p> <p style="text-align: right;">SA</p>
	<p>IT Dispositivo medico GB Medical Device FR Dispositif médical ES Producto sanitario PT Dispositivo médico DE Medizinprodukt GR Ιατροτεχνολογικό προϊόν</p> <p style="text-align: right;">SA جهاز طبي</p>
	<p>IT Dispositivo medico conforme al regolamento (UE) 2017/745 GB Medical Device compliant with Regulation (EU) 2017/745 FR Dispositif médical conforme au règlement (UE) 2017/745 ES Producto sanitario conforme con el reglamento (UE) 2017/745 PT Dispositivo médico em conformidade com a regulamento (UE) 2017/745 DE Medizinprodukt im Sinne der Verordnung (EU) 2017/745 GR Ιατρική συσκευή σύμφωνα με την ΚΑΝΟΝΙΣΜΟΣ (ΕΕ) 2017/745</p> <p style="text-align: right;">SA جهاز طبي يتوافق مع التوجيه (UE) 2017/745</p>



Gima S.p.A.
Via Marconi, 1 - 20060 Gessate (MI) Italy
gima@gimaitaly.com - export@gimaitaly.com
www.gimaitaly.com
Made in Pakistan

